



Maine Environmental Laboratory

1 Main St., Yarmouth, ME 04096 Tel.: (207) 846-6569 FAX: (207) 846-9066 E-mail: melab@mel-lab.com

Sample Containers / Preservations / Holding Times

	Solids			Aqueous		
INORGANICS	Container ¹	Preservation ²	Holding Time ³	Container ¹	Preservation ²	Holding Time ³
Metals ⁴ (except mercury)	P, G	n/a	6 months	P, G - 250 mL	HNO ₃ to pH<2	6 months
Mercury ⁴	P, G	≤6°C	28 days	P, G - 125 mL	HNO ₃ to pH<2	28 days
Alkalinity	P, G	≤6°C	n/a	P, G - 1 L	≤6°C	14 days
Ammonia-N ⁵	P, G	≤6°C	28 days	P, G - 250 mL	H ₂ SO ₄ to pH<2, ≤6°C; + pre-test kit ⁵	28 days
BOD₅ / CBOD₅	n/a	n/a	n/a	P, G - 1L	≤6°C	24 hours
Boron	P	≤6°C	6 months	P - 60 mL	HNO ₃ to pH<2	6 months
Bromide	P, G	≤6°C	n/a	P, G - 60 mL	≤6°C	28 days
Chloride	P, G	≤6°C	28 days	P, G - 60 mL	≤6°C	28 days
COD	n/a	n/a	n/a	P, G - 60 mL	H ₂ SO ₄ to pH<2, ≤6°C	28 days
Coliform (Total or Fecal), E. coli, Enterococci	n/a	n/a	n/a	P - Sterile bottle w/ Sodium Thiosulfate	DW recommended ≤10°C NPDES required ≤10°C	30 hrs-DW 8 hrs - NPDES
Color	n/a	n/a	n/a	P, G - 125 mL	≤6°C, separate bottle	24 hours
Conductivity / Specific Conductance	P, G	≤6°C	28 days	P, G - 125 mL	≤6°C	28 days
Cyanide, Total & Amenable to Chlorination	P, G	≤6°C	14 days	P, G - 100 mL	≤6°C, NaOH to pH>12	14 days
Flash Point (aqueous) Ignitability (solids)	1 - 8oz. G (fill to top)	n/a	n/a	1 - 8oz. G; fill to top	None required.	n/a
Fluoride	P	≤6°C	28 days	P - 60 mL	≤6°C	28 days
Hardness ⁴	P, G	n/a	6 months	P, G - 60 mL	HNO ₃ to pH<2 ⁶	6 months
Hexavalent Chromium	P, G	≤6°C	28 days	P, G - 500 mL	≤6°C	24 hours⁹
Legionella pneumophila	P - Sterile bottle w/ Na thiosulfate	ambient	48 hours preferred	P - Sterile bottle w/ Na thiosulfate	ambient	48 hours preferred
Nitrate-N	P, G	≤6°C	28 days	P, G - 60 mL	≤6°C	48 hours
Nitrite-N	P, G	≤6°C	28 days	P, G - 60 mL	≤6°C	48 hours
Oil & Grease	Amber G	≤6°C	n/a	1L Amber G, teflon lined cap	H ₂ SO ₄ or HCl to pH<2, ≤6°C	28 days
Orthophosphate-P & Total Reactive Phos.	n/a	n/a	n/a	P, G - 250 mL	Filter immediately, ⁶ ≤6°C	48 hours
Paint Filter	2 - 8oz G	n/a	n/a	n/a	n/a	n/a
pH	P, G	n/a	n/a	P, G - 125 mL	None required.	Immediately⁶
Radon	n/a	n/a	n/a	2 -G, no headspace ⁷	n/a	4 days
Phosphorus, Total	P, G	n/a	6 months	P, G - 250 mL	H ₂ SO ₄ to pH<2, ≤6°C	28 days
Reactivity, Cyanide	P, G	≤6°C	n/a	P, G - 200 mL	≤6°C	14 days
Reactivity, Sulfide	P, G	≤6°C	n/a	P, G - 200 mL	≤6°C	14 days

	Solids			Aqueous		
INORGANICS	Container ¹	Preservation ²	Holding Time ³	Container ¹	Preservation ²	Holding Time ³
Sulfate	P, G	≤6°C	28 days	P, G - 60 mL	≤6°C	28 days
Sulfide	P, G	≤6°C	7 days	P, G - 250 mL	ZnAcetate & NaOH to pH>9, ≤6°C	7 days
Total Dissolved Solids (TDS)	n/a	n/a	n/a	P, G - 1L	≤6°C	7 days
Total Suspended Solids (TSS)	n/a	n/a	n/a	P, G - 1L	≤6°C	7 days
Total Solids	P, G	≤6°C	7 days	P, G - 250 mL	≤6°C	7 days
Total Volatile Solids (TVS) and Loss on Ignition (LOI)	P, G	≤6°C	7 days	P, G - 250 mL	≤6°C	7 days
Tannin & Lignin	n/a	n/a	n/a	P, G - 250 mL	≤6°C	n/a
Total Kjeldahl Nitrogen (TKN)	P, G	≤6°C	28 days	P, G - 125 mL	H ₂ SO ₄ to pH<2, ≤6°C	28 days
Total Organic Carbon (TOC)	P, G	n/a	28 days	G - 40 mL (fill to top)	H ₂ SO ₄ to pH<2, ≤6°C	28 days
Turbidity	n/a	n/a	n/a	P - 250 mL	≤6°C, separate bottle	48 hours
ORGANICS						
Volatile Organic Compounds (VOCs) 524.2, 8260	G - 20 mL and total solids	methanol, ≤6°C	14 days	3 x 40 mL - glass teflon septum no headspace ⁷	3 vials w/HCl to pH<2, ≤6°C	14 days ⁸
Volatile Organic Compounds (VOCs) 624	n/a	n/a	n/a	4 x 40 mL - glass teflon septum no headspace ⁷	1 vial w/HCl to pH<2, 3 vial unpres., ≤6°C	72 hours
Semi Volatile Organic Compounds (SVOC/ABN)	Amber G	≤6°C	14 days	1L Amber G	≤6°C	7 days
MASS. VPH	G - 20 mL and total solids	methanol, ≤6°C	28 days	3 x 40 mL - glass teflon septum no headspace ⁷	HCl to pH<2 ≤6°C	14 days ⁸
MASS. EPH and TEPH (Total EPH)	4oz. Amber G	≤6°C	14 days	1L Amber G	HCl to pH<2 ≤6°C	14 days
GRO-8015 (TPH - Gas)	G - 20 mL and total solids	methanol, ≤6°C	14 days	3 x 40 mL - glass teflon septum no headspace ⁷	HCl to pH<2 ≤6°C	14 days
Total Petroleum Hydrocarbons (TPH - Diesel)	G	≤6°C	14 days	1L Amber G	≤6°C	7 days
Herbicides	Amber G	≤6°C	14 days	1L Amber G	≤6°C	7 days
PCBs	Amber G	≤6°C	14 days	1L Amber G	≤6°C	7 days
Pesticides	Amber G	≤6°C	14 days	1L Amber G	≤6°C	7 days
Total Organic Halogens (TOX)	G	≤6°C	28 days	250 mL Amber G	H ₂ SO ₄ to pH<2, ≤6°C	28 days
Methane - RSK175	n/a	n/a	n/a	3 x 40 mL - glass teflon septum no headspace ⁷	H ₂ SO ₄ to pH<2, <6°C	14 days
TCLP - METALS	P, G, 110g	If Hg, <6°C	6 mo. except Hg=28 days	P, G - 500 mL	n/a	6 mo. except Hg=28 days
TCLP - VOCs	Separate G, 4oz.	≤6°C, no headspace	14 days	4 x 40 mL - glass teflon septum no headspace If any sediment, 1 – 8oz jar filled completely	≤6°C	14 days
TCLP - SVOC, Herbicides & Pesticides	1L Amber G	≤6°C	14 days to ext., then 7 days	3 x 1L Amber Glass	≤6°C	7 days

Sample Acceptance Criteria

Sample Condition Samples will be reviewed for sample acceptance criteria upon receipt. This includes full documentation, sufficient volume, proper chemical and/or temperature preservation, appropriate sample containers and receipt of sample within hold time. If any non-compliant conditions are identified, the client will be notified for continuation or cancellation instructions. Should the client wish to proceed with analysis, the non-compliant conditions will be noted on the Chain of Custody (COC) form, the report's narrative page, and by flagging data on the final report.

Sample Integrity Shipping containers received with custody seals are checked for tampering upon receipt. Containers are checked for leaks or any conditions which might compromise the integrity of the samples. Custody seals are available from the laboratory.

Sample Documentation The laboratory provides COCs for complete documentation including sample-specific comments, and the following information: client information, sampler name, sample ID, sampling date and time, sample matrix, type of container and preservation, analytical parameters requested, and custody signatures with date and time.

Sample Labeling Sample containers should be labeled with permanent ink. Black Sharpie® works well. Include at a minimum Sample ID, date and time of sampling and any preservation included. At the lab, samples will be assigned a unique identifier matched to samples on the COC form. Water -resistant, permanent labels are available from the lab.

Sample Preservation Refer to the previous pages for specific preservation requirements. Thermal preservation 0-6° C, but not frozen.

Suggestions for Chilling Samples To ensure samples are received within the accepted range of 0-6° C (=fridge temp), use double bagged ice or **completely frozen** gel ice packs. Ice frozen in water bottles works best. Match the amount of ice at least 1:1 to the volume of sample. In cases where the sample is hand-delivered to the lab immediately after sampling, evidence that the chilling process has begun is sufficient and will be documented. When shipping a sample to the lab, waterproof coolers are recommended to maintain required temperatures. Add extra ice if the sample might sit in a UPS/Fed Ex/USPS truck or warehouse over the weekend.

Notes:

- 1 P = high density polyethylene, pre-cleaned (HDPE); G = glass, pre-cleaned. Solids "P" may use ziplock bags, if not too runny. The volumes listed may be reduced or increased depending on analyte combinations, detection limits and sample-specific quality control. Contact the laboratory for minimum volumes for specific analytical combinations.
- 2 Grab samples: immediate preservation in the field is preferred. Composite samples: preserve each aliquot at time of collection, if possible. When using an automatic sampler, cool sampler to $\leq 4^{\circ}\text{C}$ until compositing is completed.
- 3 Holding times listed are the maximum that samples may be held before analysis or extraction must begin. Holding times start at time of sampling for grab samples and end of composite period for composites.
- 4 Testing for dissolved metals requires filtration (0.45 micron) prior to preservation. Immediate filtration and preservation in the field is preferred. If this is not possible, deliver samples to the laboratory unpreserved as soon as possible for filtration and preservation. DW metals must be preserved within 14 days of sampling.
- 5 Use test kit prior to preserving in field. Request a kit and pre-test instructions with your bottle order.
- 6 EPA defines "immediately" as within 15 minutes of collection. Holding time exceedances must be noted on lab report.
- 7 Fill container(s) to top with NO HEADSPACE (no air). Invert sample to check for presence of air bubbles. Shoot for none, but 1 lentil bean-sized bubble is ok.
- 8 Holding time is 7 days when sample is not preserved with HCl to pH<2. VOC624 hold time is 72 hrs for certain compounds on list.
- 9 AQ Hexavalent chromium is subcontracted to a certified lab, with 24-hr hold time. Deliver to sub-lab directly. Holding time is 28 days if sample

QUESTIONS? Call us — we can help! 207-846-6569.